

Median machine (used with NEWARE charge and discharge equipment)		
Indicator project	Indicator parameters	
Equipment and material code:	CT-ZWJ-4'S-T-1U	
Enter the power supply	AC 220V ±10% / 50Hz	
interiorinput power	15W	
Data cache	1G B	
frequency of sampling	Max. 100Hz (10ms)	
The clock error	1S /Day	
It can be realized with the equipment	Discharge protection	Save the test state when the power is cut off for restoring the test;
	testing control	Charge and discharge the equipment according to the working process;
	real-time guard	Collect equipment data in real time, monitor the test process, and make logical protection judgment;
	off-line test	When communication with the upper computer is interrupted, ensure that the test is not affected and continue;
	Third-party device communication	Third party devices / integrated data can be controlled through communication integration (Such as: temperature box, chiller, pressure module, height scale block, press bed PLC, etc.)
CPU	ARM9	
system kernel	Linux core	
screen	LCD	
communication interface	RJ45	
communication mode	TCP/IP	
Communication rate	100M	
IP levels of protection	Protection level is IP20	
Maximum number of band channels (main channels + auxiliary channels)	256	
Equipment working environment requirements		
Indicator project	Indicator parameters	
operating temperature range	0℃~40℃	
Storage temperature range	-10℃~50℃	
Relative humidity range of the working environment	70% RH (no moisture condensation)	
Storage environment relative humidity range	80% RH (no moisture condensation)	
Case size per unit (W * D * H) (mm)	1U (19") 482*170*50	

Device pictures



Pictures are for reference only, subject to the physical object